

# Platelet PGDprime® Processing Tubes, Sample and Transfer Pipettes



## A. INTENDED USE

The Platelet PGDprime Processing Tubes, Sample and Transfer Pipettes are only for use with the Platelet PGDprime Test.

## B. SUMMARY AND EXPLANATION OF THE TEST

Platelet PGDprime Processing Tubes, Sample and Transfer Pipettes are single use, disposable accessories for use with the Platelet PGDprime Test. They are used in the processing of platelet samples prior to addition to the Platelet PGDprime Test Device and in the processing of Platelet PGDprime Controls.

## C. PRINCIPLES OF THE PROCEDURE


Platelet PGDprime Processing Tubes, Sample and Transfer Pipettes are specifically for use with the Platelet PGDprime Test. Similar laboratory disposables should not be substituted.

## D. REAGENTS AND MATERIALS

### Materials Provided

Processing Tubes, Sample and Transfer Pipettes for the Platelet PGDprime Test **REF** PRM100D-CE

Includes the following:

	100 Tests
<b>Processing Tubes</b>	100 each
<b>Sample Pipettes</b>	100 each
<b>Transfer Pipettes</b>	100 each

### Materials Required

Platelet PGDprime Test

**REF** PRM100-CE 100 Tests

Materials Required But Not Provided - Refer to the Platelet PGDprime Test Package Insert.

## E. WARNINGS AND PRECAUTIONS

For *In Vitro* Diagnostic Use

### Warnings

1. Read the Platelet PGDprime Test and Platelet PGDprime Controls package inserts completely before using the product. Follow the instructions carefully. Not doing so may result in inaccurate test results.
2. The Platelet PGDprime Processing Tubes, Sample and Transfer Pipettes have been validated for use with the Platelet PGDprime Test only. Do not substitute similar laboratory supplies.

### Handling Precautions

Handle and perform test properly: Do not re-use Processing Tubes or Pipettes.

## F. STORAGE INSTRUCTIONS

Store Platelet PGDprime Processing Tubes, Sample and Transfer Pipettes at 15 - 30 °C.

*For the following categories - Refer to the Platelet PGDprime Test Package Insert and the Platelet PGDprime Controls Package Insert.*

## G. SPECIMEN COLLECTION AND PREPARATION FOR ANALYSIS

## H. TEST PROCEDURE




## I. LIMITATIONS

## J. REFERENCES

Platelet PGDprime and Verax Biomedical are registered trademarks of Verax Biomedical Incorporated.

© 2021 Verax Biomedical Incorporated

### Keys to Symbols Used

<b>IVD</b>	<i>In vitro</i> diagnostic medical device
<b>LOT</b>	Lot Number
<b>REF</b>	List Number
<b>Sample Pipettes</b>	Sample Pipettes
<b>Transfer Pipettes</b>	Transfer Pipettes
<b>Processing Tubes</b>	Processing Tubes
	Temperature limitation
	Manufacturer
	Consult instructions for use
<b>EC REP</b>	Authorized Representative

 **VERAX BIOMEDICAL**  
148 Bartlett Street  
Marlborough, MA 01752 USA



**EC REP** **EMERGO EUROPE**  
Westervoortsedijk 60  
6827 AT Arnhem  
The Netherlands